

Olympus Winter & Ibe GmbH, Kuehnstraße 61, 22045 Hamburg

To whom it may concern

Datum

11.03.2022

## Statement from Olympus regarding unauthorized access to ENDOBASE folders

Dear Sir or Madam.

at internal security audits of ENDOBASE and ENDOVIEW it has been discovered that it is theoretically possible to bypass the ENDOVIEW user management. Under these circumstances, unauthorized read access to patient data – that should only be accessed by authorized users via ENDOVIEW – would become possible.

Olympus has developed a patch for ENDOVIEW to fix this vulnerability. The patch was distributed with service bulletin SBU\_100-236-975. All customers with affected ENDOBASE and ENDOVIEW installations were informed, and the patch was installed as agreed with the affected customers.

This vulnerability should also be removed on all remaining, unpatched installations of ENDOBASE and ENDOVIEW as soon as possible. Contact your Olympus service organization to verify if a patch is required for you system.